Technical Specification  Specification of Bed								
S.N.	Purchaser's Specification (Tamakoshi Hospital)	Yes/No	Pg. No. Catalogue	Remarks				
	Brand							
	Type/Mode							
*	Country of Origin							
	Advance Linear actuator system (04Pcs Motor, 01 Pc Control box, 01 Pc Handset, 01 PC)							
2	Four Pieces individual side-rails, ABS moulded							
3	Heavy ABS head and foot panel with locking							
l	4 section top made-up of oval perforated CRCA M.S. Sheet							
5	Framework made of M.S. Tubes	34						
3	Bed Mounted on 5" ABS Castors							
,	S.S. Telescopic I.V. Rod with 4Location							
3	Holder for Urine Bag							
)	Sitting position adjustable							
0	Pre-treated and Epoxy power coated							
1	Size: 2090L x 900W x 550-800H MM			2				
2	Position obtained by hand remote			T				
2.1	Hi-Low							
12.2	Backrest elevation							
2.3	Knee rest elevation	0						
2.4	Trendlenburg/ Reverse Trendlenburg							
3	<b>User Training</b>							
3.1	Must provide user training (including how to use and maintain the equipment).							
4	Warranty							
14.1	Comprehensive warranty for 2 year.	के रिष्ठ के	Yes					

Prof. Dr. Surian Karma NHC No. 3574 2-013

15	Maintenance Service During Warranty Period		
15.1	During the warranty period, supplier must ensure planned preventive maintenance (PPM) and corrective/breakdown maintenance whenever required.	-	ξ.
16	Installation and Commissioning	ŭ	
16.1	The bidder must arrange for the equipment to be installed and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.		
17	Documentation		
17.1	User (operating) manual in English		
17.2	Service (Technical/Maintenance) manual in English		
17.3	List of important spare parts and accessories with their part number and costing		
17.4	Certificate of calibration and inspection from factory		

The bidder must completely fill out the Technical Specification Form (TSF). Only Yes/No/All complies should not be written. The page number in the catalog of all the required parameters must be mentioned and highlighted. Failure to do so may lead to the rejection bid from the technical committee.

9

Prof. Dr. Suman Karmacharya MBBS MHCM NMC No. 3574 213